



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: May 27, 2010 INITIALS: JEC OWNER: WASHINGTON STATE PATROL
DISTRICT: 6 ADDRESS: I-90E Bullfrog
LOCATION CleElum 52 EB COUNTY: Kititas
MAKE: RiceLake TYPE Trk Scale ACCURACY CLASS: III L
MODEL: IQ-355 ID. NO: 91441423 CAPACITY: 100000 lbs. SECTIONS: 2
MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 10x10

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
CENTER	20,000	20,000	0				
SEC 1	20,000	20,000	0				
SEC 2	20,000	20,000	0				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L2902-1 and 200446-0-L2902-3

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: July 1, 2010
Print name: David Cromer

State of Washington
County of Shurston
Signed and sworn to (or affirmed before
me on July 1, 2010 by

Brenda A Biscay
(Signature of notary public)
Brenda A Biscay
Printed Name of Notary
My appointment expires 3.12.12

